

TLC Meter

Temperature, Level, Conductivity Model 107

The TLC Meter is ideal for profiling conductivity and temperature in wells and open water. It displays accurate measurements of conductivity and temperature on a convenient LCD display.

Conductivity measurements are read from 0-80,000 μ S/cm with readings giving accuracy of 5% of reading or 100 μ S (which ever is greater).

Water level and probe depth measurements are read off the Solinst durable flat tape, which is accurately laser marked each millimeter or every 1/100 ft. Tape lengths are available up to 300 m (1000 ft).

For automated water level, temperature, and conductivity datalogging, Solinst also offers the Levelogger 5 LTC. (See the Model 3001 Levelogger 5 LTC Data Sheet).

Conductivity Made Easy

- Tape lengths to 300 m (1000 ft)
- Rugged Solinst reel and accurate laser marked tape
- Probe diameter 19 mm (3/4")
- Standard 9V alkaline battery gives 90 hrs. of use
- Auto-Off after 8 minutes

Rugged Flat Tape

The high quality flat tape reels smoothly, remains flexible and hangs straight in the well, irrespective of temperature. The flat tape is mounted on a sturdy, well-balanced Solinst reel, with a convenient battery drawer for the 9 Volt alkaline battery. Permanent laser markings each millimeter or 1/100 ft allow accurate readings.

Stranded stainless steel conductors and copper coated steel conductors resist corrosion, provide strength and are nonstretch. They make the tape easy to repair and splice. The dogbone design reduces adherence to wet surfaces.





LM2: Each 1/100 ft, 1/10 ft and ft **LM3:** Each mm, cm and m

Lengths: 30 m, 60 m, 100 m, 150 m, 250 m, 300 m 100 ft, 200 ft, 300 ft, 500 ft, 750 ft, 1000 ft

[®] Solinst is registered trademark of Solinst Canada Ltd.



TLC Meter

Model 107 Data Sheet

Water Conductivity Measurements

The TLC Meter uses a 'smart' conductivity sensor with platinum electrodes to measure conductivity. The conductivity is displayed on the screen along with the associated temperature measurement. The 'smart probe' displays conductivity that has been standardized to 25°C, i.e. Specific Conductance (displayed as EC). The conductance temperature coefficient is 2.0% per °C.

Calibration is simple, using 1413 $\mu S,$ 5000 $\mu S,$ 12,880 $\mu S,$ and/or 80,000 μS solutions for 1, 2, 3, or 4 point manual conductivity calibrations.

Applications


- Profiling conductivity and temperature in wells and open water
- Salinity studies
- Saltwater intrusion investigations
- Testing for water quality impairments from road salt
- Tracer tests
- General indication of chemical contamination level
- Early warning of changes in water quality at:
 - Landfills
 - Industrial sites





Water Level & Depth Measurements

When the zero point of the probe enters water, an electric circuit is completed, briefly activating a buzzer and blanking out the screen for about 1 second. The depth to water is then read off the tape. When the TLC Meter is withdrawn from the water, a short buzz gives a warning that the probe is now out of water.



A Tape Guide/Datum allows the depth measurement to be read at the marked position on the Tape Guide, to give reliable, repeatable, accurate depth measurements when profiling a well. The Tape Guide protects the tape from damage on rough edges of well casing.

Water Temperature Measurements

The TLC Meter operates in a range from -15° C to $+50^{\circ}$ C (23°F to 122°F). You may choose to have the readings displayed on the LCD screen in °C or °F. Accuracy is +/- 0.2°C in Celsius, or +/- 0.4°F in Fahrenheit. The 'smart' TLC probe automatically adjusts the measured conductivity values to display as specific conductance standardized to 25°C. This provides standardized, repeatably comparable measurements.

When the probe is turned on, the LCD screen displays both conductivity and temperature.

Conductivity & Temperature Display Menu

The display menu is simple to operate. When the unit is turned on, it will begin to display electrical conductivity and temperature readings, in or out of water. If the button is held down for 2 seconds and released, the display moves to the next menu item. Press the button two times quickly in any given screen to achieve the displayed action.

